

Manufacturer's Declaration of Conformity

C∈ marking in accordance with the Medical Device Regulation (EU) 2017/745

| Manufacturer's name: Manufacturer's Adress: | Lopital Nederland B.V. Laarakkerweg 9, 5061 JR, Oisterwij The Netherlands | k | |
|---|---|-----------------------------------|--|
| Manufacturer's SRN (Single Registration Number): | NL-MF-000004372 | | |
| Brand Name: | Lopital | | |
| Medical device: Model number(s): Device Description: | Sirocco basic & Sirocco deluxe 61002500 & 61002510 Wall mounted Shower Stretcher wi adjustment | th electrically powered high/ low | |
| Basic UDI-DI: | 872025610356161002500L7 & 872 | 2025610357861002510S8 | |
| Classification: | Class I | | |
| Conforms to regulation: | Medical Device Regulation (EU) 2017/745 | | |
| Standards applied: | NEN-EN-ISO 14971:2019 | | |
| The product has been tested and evaluated in accordance with the following standards: | ISO 17966:2016 EN 60601-1-2 (2015)+A1(2021) NEN-EN-IEC 62366-1:2015 | | |
| The product has been designed and manufactured under a certified quality management system in accordance with: | UNI-CEI-EN-ISO 13485:2016 | | |
| This declaration of conformity is issued under the sole responsibility of Lopital. We hereby declare that the medical device(s) specified above meet the provision of the Regulation (EU) MDR 2017/745 for medical devices. | | | |
| Signature: | Date: | Place: | |

Oisterwijk

Jan Van Megen, CEO dd-mm-yyyy

12-12-2024

Document Number: DOC2500-2024-03